Nelson Labs™
A Sotera Health company

Sponsor:
Wendover Brown
Ohlone Press, LLC
1830 Leavenworth St.
San Francisco, CA 94109

Bacterial Filtration Efficiency (BFE) Final Report

Test Article:	95F - 1
Purchase Order:	101818
Study Number:	1112599-S01
Study Received Date:	23 Oct 2018
Testing Facility:	Nelson Laboratories, LLC 6280 S. Redwood Rd. Salt Lake City, UT 84123 U.S.A.
Test Procedure(s):	Standard Test Protocol (STP) Number: STP0004 Rev 15
Deviation(s):	None

Summary: The BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of *Staphylococcus aureus* was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at 1.7 - 2.7 x 10^3 colony forming units (CFU) with a mean particle size (MPS) of 3.0 ± 0.3 µm. The aerosols were drawn through a six-stage, viable particle, Andersen sampler for collection. This test method complies with ASTM F2101-14, EN 14683:2014, Annex B, and AS4381:2015.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side:	Either
BFE Test Area:	~40 cm^2
BFE Flow Rate:	28.3 Liters per minute (L/min)
Conditioning Parameters:	85 ± 5% relative humidity (RH) and 21 ± 5°C for a minimum of 4 hours
Positive Control Average:	2.5 x 10^3 CFU
Negative Monitor Count:	<1 CFU
MPS:	3.0 µm

ilac-MRA ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY

Janelle Bentz
Study Director Janelle R. Bentz, M.S.

31 Oct 2018
Study Completion Date

1112599-S01

Results:

Test Article Number	Percent BFE (%)
1	>99.9[a]
2	>99.9[a]
3	>99.9[a]
4	>99.9[a]
5	>99.9[a]

[a] There were no detected colonies on any of the Andersen sampler plates for this test article.

The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} x\ 100$$

C = Positive control average
T = Plate count total recovered downstream of the test article
Note: The plate count total is available upon request